



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALIKINARU JN., TIRUNELVELI DISTRICT – 627105

Date: 05.05.2023

MINUTES OF EQUIPMENT MEETING REPORT

Held in Principal Chamber on 05.05.2023 at 9.00 A.M

Prosthodontics :

1. Motor driven Centrifugal casting machine – Taken by the electrical department for the conversion of manual spring loaded casting machine to motorize casting machine.
2. Hydro solder (welder) to be got from Bharath Dental Lab.
2. Steam cleanser nozzle washer needs to be changed. - waiting for spare
3. Dewaxing Unit-3No.digital display control Unit- Panel to be changed.- Pending for AMC
4. Hot Water sterilizer coil has to be Repaired- MRL given.
5. Acrylizer coil- 4 Nos.has be Repaired- MRL given.
6. Burnout Furnance – 2Nos Not Working.
7. Ceramic furnace – 1No. Not working (from Conservative Dentistry dept).

PG:

1. Circuit issue seen in the both plug points of Acrylizing unit and sterilization unit.

Public Health Dentistry:

1. BP Apparatus one more to be purchased before DCI inspections- Requested to Purchase Dept
2. Mobile dental unit AC need to be repaired – General Maintenance

Oral Pathology

1. Digital Microscope Display is not Working - Payment made – Display not repaired yet. Supplier to be contacted.
2. Water Bath Slide warming table to be upgraded.- Requested to Purchase Dept



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALIKINARU JN - 627 105.
TIRUNELVELI DISTRICT.

Page 1 of 4



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai)
Accredited by NAAC
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

3. Chair no 10, 11 compressor need to repair.
4. Slides box and slides staining rack need.
5. Monocular light Microscope charge is draining and not working.
Solutions: 1) Under the slab plug points (provision needed) can be installed.
2) Need to change battery.

Periodontics

UG

1. B class Auto clave purchased not been installed.
2. Chair no.3 & 13 light is not working (It is repairing by Mr.Raja)
3. 2 chairs urgently required for inspection purpose.

Conservative Dentistry

UG

1. Need 2 apex locator - Requested to Purchase Dept
2. Required 1 endomotor - Requested to Purchase Dept
3. Automatic developer not working- Need to be condemned
4. Indent for Auto clave given on 22.6.2021 .Urgently needed for inspection purpose.
5. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- General Maintenance
6. Induction casting machine not working- waiting for spares
7. Computer in RVG room not working - General Maintenance
8. Amalgamator : Needs pre proportioned capsules - 3 sets
9. Carborundum disc and diamond disc required.

PG

1. Endomotor required.- Requested to Purchase Dept. Only one needed endomotor repaired by Mr.Raja.
2. Pulp tester not working and also need new one.
3. Required 1 endomotor - Requested to Purchase Dept
4. Need 2 apex locator - Requested to Purchase Dept
5. Bleaching unit needed



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

6. Putty material extruder required
7. IOPAR X ray Machine PCB board to Bangalore for repair on 21.02.2023.
8. RVG/PSP plates for department. Current RVG declared as condemned by satelec sopix company representatives.
9. Vita 1 line ceramic furnace motherboard complaint not rectified.

Pedodontics

UG

1. Phantom Heads 5 Units Installation and Maintenance – spare parts to be bought from Dhumraketu - waiting for spares

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement - Requested to Purchase Dept
2. Need one scaler unit in the UG- Requested to Purchase Dept. Urgently required for inspection purpose.

UG

1. Operator chair - 8Nos required.- Requested to Purchase Dept

Oral Surgery

1. Rhinoplasty kit required- Request given to Purchase Manger.

UG

1. Compressor Airline to be given to all chairs. - waiting for spares.

Oral Medicine

1. Automatic developer machine need to be repair.
2. Automatic developer solution needs to be procured.



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Suggestion

1. PSP and hand held X ray Portable Unit quotation given.

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:-

Dr. Amalorpavam-Department of Prosthodontics

Dr. John Hearty Deepak, Department of Oral Medicine

Dr. Esai Amutha Prabha- Department of Pedodontics

Dr. Divakar, Department of Oral Surgery

Dr. Jevina Christy, Department of Conservative Dentistry

Dr. M.S. Ravishankar, Department of Conservative Dentistry

Dr. Ani Simila, Department of Oral Pathology

Dr. P. Fairlin, Department of Periodontics

Dr. C. Selvakumar, Department Of Public Health Dentistry

Dr. Bevin Shaga Department of Orthodontics



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.

Date: 24.03.2023

MINUTES OF EQUIPMENT MEETING REPORT

Held in Principal Chamber on 24.03.2023 at 9.00 A.M

Prosthodontics :

1. Motor driven Centrifugal casting machine - Mr.Subbiah to follow up.
2. Steam cleanser nozzle washer needs to be changed. - waiting for spare
3. Dewaxing Unit-3No.digital display control Unit- Panel to be changed.- Pending for AMC
4. Hot Water sterilizer coil has to be Repaired- MRL given.
5. Acrylizer coil- 4 Nos.has be Repaired- MRL given.
6. Burnout Furnance – 2Nos Not Working.
7. Ceramic furnace – 1No. Not Working (from Conservative Dentistry dept)

Public Health Dentistry:

1. BP Apparatus one more to be purchased before DCI inspections- Requested to Purchase Dept
2. Mobile dental unit AC need to be repaired – General Maintenance
3. Aralvaimozhi Rural Satellite clinic 2 Dental Chair & Compressor installed. Interior work in process, Racks and Table need to be installed. Name Board wanted.

Oral Pathology

1. Digital Microscope Display is not Working - Payment made – Display not repaired yet. Supplier to be contacted.
2. Water Bath Slide warming table to be upgraded.- Requested to Purchase Dept
3. Chair no 11, 12 compressor need repair.
4. Slides box and slides staining rack need.
5. Monocular light Microscope charge is draining and not working.
Solutions: 1) Under the slab plug points (provision needed) can be installed.
2) Need to change battery.



A handwritten signature in green ink, appearing to be "D. Subbiah".

Periodontics

UG & PG

1. B class Autoclave purchased not been installed.
2. 2 dental chair + 2 scaler unit + 3 tips + 1 chuck shifted to Aralvaimozhi outpost.

Conservative Dentistry

UG

1. Need 2 apex locator - Requested to Purchase Dept
2. Required 1 endomotor - Requested to Purchase Dept
3. Automatic developer not working- Need to be condemned
4. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- General Maintenance
5. Induction casting machine not working- waiting for spares
6. Computer in RVG room not working - General Maintenance
7. Amalgamator : Needs pre proportioned capsules - 3 sets
8. Carborundum disc and diamond disc required.

PG

1. Endomotor required.- Requested to Purchase Dep
2. Pulp tester not working
3. Required 1 endomotor - Requested to Purchase Dept
4. Need 2 apex locator - Requested to Purchase Dept
5. Bleaching unit needed
6. Putty material extruder required
7. IOPAR X ray Machine PCB board to Bangalore for repair on 21.02.2023.
8. RVG/PSP plates for department. Current RVG declared as condemned by satelec sopix company representatives.
9. Vita 1 line ceramic furnace motherboard complaint not rectified.



A handwritten signature in green ink, appearing to be "Dr. [Name]".

Pedodontics

UG

1. Phantom Heads 5 Units Installation and Maintenance – spare parts to be bought from Dhumraketu - waiting for spares

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement - Requested to Purchase Dept
2. Need one scaler unit in the UG- Requested to Purchase Dept

UG

1. Operator chair - 8Nos required.- Requested to Purchase Dept

Oral Surgery

1. Rhinoplasty kit required- Request given to Purchase Manger.

UG

1. Compressor Airline to be given to all chairs. - waiting for spares.

Oral Medicine

1. OPG Machine no clarity- quotation is being taken for buy back and replacement.- waiting for quotation approval
2. IOPA Machine installation was done and pending for an approval

Suggestion

1. PSP and hand held X ray Portable Unit quotation given.



A handwritten signature in green ink, appearing to be "Dinesh".

PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:-

- Dr.Amalorpavam-Department of Prosthodontics
- Dr.John Hearty Deepak, Department of Oral Medicine
- Dr.Esai Amutha Prabha- Department of Pedodontics
- Dr.Divakar, Department of Oral Surgery
- Dr.Jevina Christy, Department of Conservative Dentistry
- Dr. Ani Simila, Department of Oral Pathology
- Dr.M.Thamilselvan, Department of Periodontics
- Dr.C.Selva kumar,Department Of Public Health Dentistry
- Dr. Bevin Shaga Department of Orthodontics

PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.





RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai)
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Date: 28.06.2023

MINUTES OF EQUIPMENT MEETING REPORT

Held in Principal Chamber on 28.06.2023 at 9.00 A.M

Prosthodontics :

1. Steam cleanser nozzle washer needs to be changed. - waiting for spare
2. Digital dewaxing unit needs to be verified from manufacturers. Requested for AMC service.
3. Digital acrylizer unit needs to be verified from manufacturers. Requested for AMC service.
4. Burnout Furnance – i) Operating circuit issue ii) Not closing properly

Public Health Dentistry:

1. Mobile dental unit AC need to be repaired – General Maintenance
2. Compressor-1hp need to be change in Aralvaimozhi rural satellite clinic

Oral Pathology

1. Digital Microscope Display is not Working -Complaint from January 2023
 - Approval pending in February 2023.
 - March 2023 proposal sent & quotation received.
 - April 2023 Mr.Ligesh reported as payment made.
 - May 2023 Maintenance department reported as spares are not available and payment not made.
 - June 2023 Accepted for replacement but still not received.
 - 8th June 2023 taken for repair by Lawrence & Mayo
 - Quotation to be sent by company. Email sent to company.





PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

2. Water Bath Slide warming table to be upgraded.- Requested to Purchase Dept
3. Chair no 1 –Front & back chair control is not working. Hand rest broken.
4. Chair no 10, 11 compressor need to repair.
5. Slides box and slides staining rack need.
6. Monocular light Microscope charge is draining and not working.
Solutions: 1) Under the slab plug points (provision needed) can be installed.
2) Need to change battery.

Periodontics

PG & UG

Nil

Conservative Dentistry

UG

1. Need 2 apex locator - Requested to Purchase Dept
2. Required 1 endomotor - Requested to Purchase Dept
3. Automatic developer not working- Need to be condemned
4. Indent for Auto clave given on 22.6.2021 .
5. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- General Maintenance
6. Computer in RVG room not working - General Maintenance
7. Amalgamator : Needs pre proportioned capsules - 3 sets
8. Carborundum disc and diamond disc required.

PG

1. Endomotor not working- Given for repair.
2. Pulp tester not working and also need new one.
3. Need 2 apex locator - Requested to Purchase Dept
4. Putty material extruder required
5. IOPAR X ray Machine PCB board to Bangalore for repair on 21.02.2023.
6. RVG/PSP plates for department. Current RVG declared as condemned by satelec sopix company representatives.
7. Vita 1 line ceramic furnace motherboard complaint not rectified.





PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Pedodontics

UG

1. Phantom Heads 5 Units Installation and Maintenance – spare parts to be bought from Dhumraketu - waiting for spares

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement - Requested to Purchase Dept
2. Need one scaler unit in the UG- Requested to Purchase Dept. Urgently required for inspection purpose.

UG

1. Operator chair - 8Nos required.- Requested to Purchase Dept
2. Double disc model trimmer under maintenance – 1 Disc need to be replaced
3. Department printer is not working- Taken it for repair

Oral Surgery

1. Rhinoplasty kit required- Request given to Purchase Manger.
2. Auto clave is repair in Operation Theatre.

UG

1. Compressor Airline to be given to all chairs. - waiting for spares.

Oral Medicine

1. IOPA machine is not functioning – Alerio DC optoma mobile unit




PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105. Page 3 of 4
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:-

Dr. Amalorpavam-Department of Prosthodontics

Dr. John Hearty Deepak, Department of Oral Medicine

Dr. Esai Amutha Prabha- Department of Pedodontics

Dr. Divakar, Department of Oral Surgery

Dr. K. T. Manoj, Department of Conservative Dentistry

Dr. Ani Simila, Department of Oral Pathology

Dr. M. Thamilselvan, Department of Periodontics

Dr. P. Fairlin, Department of Periodontics

Dr. C. Selvakumar, Department Of Public Health Dentistry

Dr. Bevin Shaga Department of Orthodontics

PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.





RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)

Kavalkinaru Jn. Tirunelveli District - 627105

Date: 08.11.2023

Agenda:

- Discussion about submitting dental chair report
- Discussion about chair side assistant duty
- Discussion about proper washing of booster bottles
- Discussion about requisition for equipments and consumables

Minutes of Equipment Committee Members Meeting

Held in Principal Chamber on 08.11.2023 at 9 AM

- The Principal instructed all equipment committee in charges of each department to prepare a dental chair report which includes the chair number, working condition of all systems and other remarks with concerned in charge signature. And submit report for both UG and PG section.
- The Principal explained the importance of dental chair report, he stated that Oral Medicine and Orthodontics department has old dental chairs; we are planning to replace it and purchase new dental chairs for the respective departments. So if every department pro actively gives the baseline data of dental chair report means it is helpful to know about which are the departments which need replacements of dental chairs.
- The Principal advised to allot a dental chair for each PGs and as well as UGs during their posting in each department because some of the chairs are not in use regularly. If they will use regularly means we will come to know about the working condition of the chair and keep the record of it while preparing equipment report every month.
- The Principal instructed that chair side assistants have to assist the staff and PGs while they are performing clinical work, but those sisters are not assisting them, so the Principal advised staff members to instruct chair side assistant to assist while treatment is given.
- The Principal stated that chair side assistants are simply sitting in the reception area and using mobiles phones. And staff members also do the same and even they not aware about arrival of the Principal to their table,



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT,



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

this nature should change and staffs should be pro actively involved in the department clinical works.

- The Principal stated that trainee sisters are allotted to each department but they are not working properly, he advised staff members to instruct them to clean the booster bottles every day and also clean the dental chairs also, because most of the department booster bottles are filthy. And also instructed to use filter water to fill the bottles. Trainee sister should make a routine for cleaning of dental chair, booster bottles and disinfection of chairs as per chair no (1-10 on day1, 11-20 on day 2 etc).
- The Principal stated that all trainee sister are going to lunch together, this should be avoided and they must be taking working lunch time for 15mins only on rotational basis only between 12.30 to 1.30 PM and this should be strictly monitored by staff and concerned section in charge of each department.
- The Principal advised staff members to bring requisition for equipments and consumables such as bottle cleaning brush, disinfectant and disposable cups.

Other Discussions:

- Dr. John Hearty Deepak stated that the intensity of chair in Biodent light is low even in high intensity option, so it needs to be rectified. The Principal replied that it is not possible to change the light in bulk on every chair; only we can change a fused light.
- Dr. Bevin Shaga stated that, during the middle of procedure biodent chair stops working frequently. The Principal asked him to give a report according that we will rectify it as soon as possible.

Meeting Chaired By : Prof. Dr. Alex Mathews Muruppel

Meeting Attended By : Dr. J. John Hearty Deepak

Dr. Akash Mithran

Dr. C. Chrisantha Joybell



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Dr.A.Esai Amutha Prabha

Dr.K.T.Manoj

Dr.Anu Krishnakumar

Dr.M.Thamilselvan

Dr.P.Fairlin

Dr.I.Bevin Shaga

Dr.C.S.Ani Simila

Dr.T.K.Divakar

Dr.Kala Bagavathy



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Date: 25.08.2022

MINUTES OF EQUIPMENT REPORT

Held in Principal Chamber on 25.08.2022 at 9.30 A.M.

Prosthodontics

1. Ray Foster high speed grinder to be checked on priority on visit of AMC person from Odisha
2. Centrifugal casting machine with motor attachment has to be followed up Mr.Subbiah

Public Health Dentistry

1. BP Apparatus one more to be purchased before DCI inspections
2. Mobile dental unit AC need to be repaired
3. Radhapuram Satellite clinic to be shifted to Aralvailmozhi – Permission sought

Oral Pathology

1. pH meter is not working
2. Need two microscope for PG

Periodontics

UG

1. Chair no. 24 and 25 – light is not working – **on priority with Ms.Preethika**
2. Need Autoclave 20 litre capacity – **Periodic Inspection Requirement**

PG

1. Chair no. 6 X-ray viewer not working
2. UNC 15 probe – 3 not available

Conservative Dentistry

UG

1. Need 21 units of phantom heads to be installed in phantom heads lab – **Confirm order and delivery with purchase department**
2. 20 mouth mirrors urgently required (Order made for 30 mouth mirror in March 2021) – Received 20 nos in March 2021, Present order given last month
3. Need 3 (nos) B class autoclave with 21 litre capacity - **Periodic Inspection Requirement**
4. Need 2 apex locator
5. Required 1 endomotor
6. Automatic developer not working
7. Replacement of cupboards in Sterilization room - indent given on 22.04.2021
8. Induction casting machine not working
9. Pulp tester not working
10. Computer in RVG room not working
11. Amalgamator : Needs pre proportioned capsules - 3 sets



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

PG

1. Ceramic Furnace motherboard not working - **on priority with Ms.Preethika**
2. RVG sensor is not working- Old RVG sensor is given for repair, company representative has returned back that RVG cannot be repaired
3. X-Ray machine unit under repair (AMS)

Pedodontics

1. No issues

Orthodontics

1. 2 Ultra sonic cleaner required **for UG as per DCI requirement**
2. Pressure moulding device need service
3. Plaster dispenser motor not working has to be repaired before DCI inspection
4. Need one scaler unit

Oral Surgery


1. Need 10 mouth mirrors and 10 periosteal elevators – **indent submitted**

Oral Medicine


1. No issues


Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel 

Meeting attended by:- - Dr.Divakar, Department of Oral Surgery

- Dr.Sarathchandra Govindraj, Department of Prosthodontics 


- Dr.Sudharani, Department of Oral Pathology

- Dr.Bevin Shaga, Department of Orthodontics 

- Dr.P.Fairlin, Department of Periodontics 

- Dr.M.Thamilselvan, Department of Periodontics 

- Dr.S.Pallavi, Department of Conservative Dentistry

- Dr.Vinej Somaraj, Department of Public Health Dentistry 

- Dr.Chrishantha Joybell, Department of Pedodontics 



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Date: 28.10.2022

MINUTES OF EQUIPMENT REPORT

Held in Principal Chamber on 28.10.2022 at 12.00 P.M

Prosthodontics

1. Motor driven Centrifugal casting machine - Mr.Subbiah to follow up.
2. Steam cleanser nozzle washer needs to be changed.

Public Health Dentistry

1. BP Apparatus one more to be purchased before DCI inspections
2. Mobile dental unit AC need to be repaired
3. Radhapuram Satellite clinic to be shifted to Aralvailmozhi – Permission sought

Oral Pathology

1. No issues

Periodontics

UG

1. Chair no. 24 and 25 – light is not working
2. Need Autoclave 20 litre capacity – **Periodic Inspection Requirement**
3. Chair no: 16 light is not working – MRL Given on 29/09/2022

PG

1. Chair no.6 X-ray viewer not working

Conservative Dentistry

UG

1. Need 2 apex locator
2. Required 1 endomotor
3. Automatic developer not working
4. Replacement of cupboards in Sterilization room - indent given on 22.04.2021
5. Induction casting machine not working
6. Pulp tester not working
7. Computer in RVG room not working
8. Amalgamator : Needs pre proportioned capsules - 3 sets

PG

1. Hand piece cleanser wall mounting need to be done
2. Endo motor servicing need to be done



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Pedodontics

1. No issues

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement
2. Need one scaler unit in the UG

Oral Surgery


1. No issues

Oral Medicine


1. No issues

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel 

Meeting attended by:- - Dr.GnanaSagar, Department of Oral Surgery

- Dr.Akash Mithran, Department of Prosthodontics 

- Dr.Sudharani, Department of Oral Pathology


- Dr.Bevin Shaga, Department of Orthodontics 

- Dr.P.Fairlin, Department of Periodontics

- Dr.M.Thamilselvan, Department of Periodontics 

- Dr.S.Pallavi, Department of Conservative Dentistry

-Dr.M.S.Ravi sankar,Department Of Conservative Dentistry

- Dr.Vinej Somaraj, Department of Public Health Dentistry 

- Dr.Chrishantha Joybell, Department of Pedodontics

-Dr.Nelson,Department of Oral Medicine



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Date: 28.11.2022

MINUTES OF EQUIPMENT REPORT

Held in Principal Chamber on 28.11.2022 at 12.00 P.M

Prosthodontics

1. Motor driven Centrifugal casting machine - Mr.Subbiah to follow up.
2. Steam cleanser nozzle washer needs to be changed.
3. Dewaxing Unit-3No.digital display control Unit- Panel to be changed.
4. Hot Water sterilizer coil has to be Repaired- MRL given.
5. Acrylizer coil- 4 Nos.has be Repaired- MRL given.
6. Vibrator- 4Nos.to be Repaired-MRL given.

Public Health Dentistry

1. BP Apparatus one more to be purchased before DCI inspections
2. Mobile dental unit AC need to be repaired
3. Radhapuram Satellite clinic to be shifted to Aralvaimozhi – Permission sought

Oral Pathology

1. No issues

Periodontics

UG

1. Chair no: 24 and 25 – light is not working
2. Need Autoclave 20 litre capacity – **Periodic Inspection Requirement**
3. Chair no: 16 light is not working – MRL Given on 29/09/2022

PG

1. Chair no: 6 -Airotor slot water leakage
2. Chair no: 7- Lights not Working.


MRL Submitted

1. Hand Control-Not working-Chair no.18,23,27,32 - 4Nos
2. Foot Control-Not working- Chair no.1,2,3,4,8,7,11,12,14,18,19,22,23,26,27- 14Nos
3. Light Button- Not working- Chair no.16,25- 2Nos
4. Chair loose - Chair no.3,15,16,21,29- 5Nos
5. Light loss- Chair no.6,7,10,11,12,21,23- 7Nos
6. Light Not working- Chair no.24,25,26- 3Nos
7. Spittoon water Not working- Chair no.6,7,8,24,26- 5Nos

Conservative Dentistry

UG

1. Need 2 apex locator
2. Required 1 endomotor



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

3. Automatic developer not working
4. Replacement of cupboards in Sterilization room - indent given on 22.04.2021
5. Induction casting machine not working
6. Pulp tester not working
7. Computer in RVG room not working
8. Amalgamator : Needs pre proportioned capsules - 3 sets
9. Preclinical lab no.2, Chair no.5 phantom head torso broken on 28.11.2022.
10. IOPAR Machine (PG) needs repair, to be sent to Bangalore.

PG

Pedodontics

1. No issues

Orthodontics

1. 2 Ultra sonic cleaner required **for UG as per DCI requirement**
2. Need one scaler unit in the UG

Oral Surgery

1. Rhinoplasty kit required- request given to Purchase Manger.

Oral Medicine

1. OPG Machine no clarity- quotation is being taken for buy back and replacement.

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Murupel

Meeting attended by:-

- Dr.Angelin Deepthi, HOD-Department of Oral Medicine
- Dr.Aarti Rajambigai, HOD-Department of Prosthodontics
- Dr.S.D.Milling Tania, HOD-Department of Orthodontics
- Dr.Chrishantha Joybell,HOD- Department of Pedodontics
- Dr.Divakar, Department of Oral Surgery
- Dr.Sarathchandra Govind Raj, Department of Prosthodontics
- Dr.Sudha Rani, Department of Oral Pathology
- Dr.Bevin Shaga, Department of Orthodontics
- Dr.P.Fairlin, Department of Periodontics
- Dr.M.Thamilselvan, Department of Periodontics
- Dr.S.Pallavi, Department of Conservative Dentistry
- Dr.C.Selva kumar,Department Of Public Health Dentistry



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Date: 30.01.2023

ACTION TAKEN REPORT

Regarding maintenance committee meeting Held in Principal Chamber on 30.01.2023 at 10.30 A.M

Prosthodontics

1. Motor driven Centrifugal casting machine - **Mr.Subbiah to follow up.**
2. Steam cleanser nozzle washer needs to be changed. - **waiting for spare**
3. Dewaxing Unit-3No.digital display control Unit- Panel to be changed.- **Pending for AMC**
4. Hot Water sterilizer coil has to be Repaired- MRL given.- **Work Finished**
5. Acrylizer coil- 4 Nos.has be Repaired- MRL given.- **Work Finished**
6. Micromotor Hand Piece - 1Nos. the card and Management Push required under Purchase. - **Work Finished**

Public Health Dentistry

1. BP Apparatus one more to be purchased before DCI inspections- **Requested to Purchase Dept**
2. Mobile dental unit AC need to be repaired – **General Maintains**
3. Radhapuram Satellite clinic to be shifted to Aralvaimozhi – Permission sought – **Working Progress**

Oral Pathology

1. Digital Microscope Display is not Working.- **Approval Pending – Display**
2. WaterBath Slide warming table to be upgraded.- **Requested to Purchase Dept**

Periodontics

UG


1. Ultrasonic scalers – 4,9,11 not working. - **Work Finished**
2. Need Autoclave 20 litre capacity – **Periodic Inspection Requirement**
3. Spitoon tumbler 13 Nos not working. - **Work Finished**

PG

1. Auto clave purchased not been installed.- **Work Finished**
2. Light cure units adapter repair. - **Work Finished**

MRL Submitted

1. Hand Control-Not working-Chair no.18,23,27,32 - 4Nos- **Work Finished**
2. Foot Control-Not working- Chair no.1,2,3,4,8,7,11,12,14,18,19,22,23,26,27- 14Nos- **Work Finished**
3. Light Button- Not working- Chair no.16,25- 2Nos - **Work Finished**
4. Chair loose - Chair no.3,15,16,21,29- 5Nos - **Work Finished**
5. Light loss- Chair no.6,7,10,11,12,21,23- 7Nos - **Work Finished**



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

6. Light Not working- Chair no.24,25,26- 3Nos- **Work Finished**
7. Spittoon water Not working- Chair no.6,7,8,24,26- 5Nos - **Work Finished**

Conservative Dentistry

UG

1. Need 2 apex locator - **Requested to Purchase Dept**
2. Required 1 endomotor - **Requested to Purchase Dept**
3. Automatic developer not working- **Need to be condemned**
4. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- **General Maintenance**
5. Induction casting machine not working- **waiting for spare**
6. Pulp tester not working- **Work Finished**
7. Computer in RVG room not working - **General Maintenance**
8. Amalgamator : Needs pre proportioned capsules - 3 sets - **store**
9. Preclinical lab no.2, Chair no.5 phantom head torso broken on 28.11.2022. - **Work Finished**
10. IOPAR Machine (PG) needs repair, to be sent to Bangalore.- **waiting for spare**
11. Light cure units adapter repair. - **Work Finished**

PG

1. Endomotor required.- **Requested to Purchase Dep**
2. light Cure units not working.- **Work Finished**

Pedodontics

UG

1. Phantom Heads 5 Units Installation and Maintance – spair parts to be bought from Dhumriketu - **waiting for spare**

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement - **Requested to Purchase Dept**
2. Need one scaler unit in the UG- **Requested to Purchase Dept**

UG

1. Operator chair - 8Nos required.- **Requested to Purchase Dept**



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Oral Surgery

1. Rhinoplasty kit required- **Request given to Purchase Manger.**

UG

2. Compressor Airline to be given to all chairs.- **waiting for spare**

Oral Medicine

1. OPG Machine no clarity- quotation is being taken for buy back and replacement.- **waiting for quotation approval**
2. IOPA Machine not working - Board Problem MRL given.- **waiting for spare**
3. Chair No – 1,4,9 Up and down movement switch missing.- **Work Finished**

PG

1. Chair no- 5 Light is not working.- **Work Finished**

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:-

-Dr.Amalorpavam-Department of Prosthodontics

-Dr.Nelson, Department of Oral Medicine

- Dr.Esai Amutha Prabha- Department of Pedodontics

- Dr.Divakar, Department of Oral Surgery

- Dr.Jevina Christy, Department of Conservative Dentistry

- Dr. Ani Simila, Department of Oral Pathology

- Dr.Bevin Shaga, Department of Orthodontics

- Dr.P.Fairlin, Department of Periodontics

- Dr.M.Thamilselvan, Department of Periodontics

- Dr.Manoj, Department of Conservative Dentistry

-Dr.C.Selva kumar,Department Of Public Health Dentistry



(Handwritten signature)

PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Date: 30.09.2022

MINUTES OF EQUIPMENT REPORT

Held in Principal Chamber on 30.09.2022 at 9.30 A.M.

Prosthodontics

1. Ray Foster high speed grinder – has to be procured from IDS Delhi – Specification has to be given by AMC person.
2. Motor driven Centrifugal casting machine - Mr.Subbiah to follow up.
3. Steam cleanser nozzle washer needs to be changed.

Public Health Dentistry

1. BP Apparatus one more to be purchased before DCI inspections
2. Mobile dental unit AC need to be repaired
3. Radhapuram Satellite clinic to be shifted to Aralvailmozhi – Permission sought

Oral Pathology

1. No issues

Periodontics

UG

1. Chair no. 24 and 25 – light is not working – **on priority with Ms.Preethika**
2. Need Autoclave 20 litre capacity – **Periodic Inspection Requirement**
3. Chair no: 16 light is not working – MRL Given on 29/09/2022

PG

1. Chair no. 6 X-ray viewer not working
2. UNC 15 probe – 3 not available to be purchased.

Conservative Dentistry

UG

1. Need 2 apex locator
2. Required 1 endomotor
3. Automatic developer not working
4. Replacement of cupboards in Sterilization room - indent given on 22.04.2021
5. Induction casting machine not working
6. Pulp tester not working
7. Computer in RVG room not working
8. Amalgamator : Needs pre proportioned capsules - 3 sets

PG

1. New RVG sensor required.



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R. Medical University, Accredited by NAAC)
Kavalkinaru Jn. Tirunelveli District - 627105

Pedodontics

1. No issues

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement
2. Pressure moulding device Handle .
3. to be made by Mr.Arumugaraja.
4. Need one scaler unit in the UG

Oral Surgery

1. No issues

Oral Medicine

1. No issues

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:- - Dr.Divakar, Department of Oral Surgery

- Dr.Sarathchandra Govindraj, Department of Prosthodontics

- Dr.Sudharani, Department of Oral Pathology

- Dr.Bevin Shaga, Department of Orthodontics

- Dr.P.Fairlin, Department of Periodontics

- Dr.M.Thamilselvan, Department of Periodontics

- Dr.S.Pallavi, Department of Conservative Dentistry

- Dr.Vinej Somaraj, Department of Public Health Dentistry

- Dr.Chrishantha Joybell, Department of Pedodontics

-Dr.Nelson,Department of Oral Madicine



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Date: 25.08.2023

MINUTES OF EQUIPMENT MEETING REPORT

Held in Principal Chamber on 25.08.2023 at 9.40 A.M

Prosthodontics :

1. Steam cleanser nozzle washer needs to be changed. - waiting for spare
2. Digital dewaxing unit needs to be verified from manufacturers. Requested for AMC service.
3. Digital acrylizer unit needs to be verified from manufacturers. Requested for AMC service.
4. Burnout Furnance – i) Operating circuit issue ii) Not closing properly
5. Ray Foster spindle grinder second unit also not working (total of 2 units only). Both are non functional. – Spindle was couriered to the IDS Chennai on 08.08.2023.
6. UG clinic: **Staff table repair** - MRL given on 21.07.2023

Public Health Dentistry:

1. **Mobile dental unit AC need to be repaired** – General Maintenance.
2. Compressor-2hp need to be installed in Aralvaimozhi rural satellite clinic.

Oral Pathology

1. Digital Microscope Display is not Working -Complaint from January 2023
 - Approval pending in February 2023.
 - March 2023 proposal sent & quotation received.
 - April 2023 Mr.Ligesh reported as payment made.
 - May 2023 Maintenance department reported as spares are not available and payment not made.
 - June 2023 Accepted for replacement but still not received.



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

- 8th June 2023 taken for repair by Lawrence & Mayo
 - Quotation was requested to company. Email sent to company on 28.06.2023. Quotation sent by company on 01.07.2023.
 - Irreparable old display was sent to us by courier.
 - For new display quotation has been given to COO on 18.07.2023.
 - 14.08.2023 demo for new digital display shown in the department in the presence of COO.
2. Water Bath Slide warming table to be upgraded.- Requested to Purchase Dept – Requisition given on 28.07.2023
 3. Chair no 10, 11 compressor need to repaired – Request for oil free compressor on 28.07.2023.
 4. Slides box and slides staining rack needed- Requisition given on 28.07.2023.
 5. Monocular light Microscope charge is draining and not working.
Solutions: 1) Under the slab plug points (provision needed) can be installed.
2) Need to change battery. – URGENT needed for practical exam.
 6. Printer is not working, needs to be repaired.

Periodontics

PG :

Nil

UG:

1. In scaler 9 & 11 increase amount of water is flowing while using.
2. Chair no 10 & 25 light is not working.

Conservative Dentistry

UG

1. Need 2 apex locator - Requested to Purchase Dept


PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.

Page 2 of 4





RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

2. Required 1 endomotor - Requested to Purchase Dept
3. Automatic developer not working- Need to be condemned
4. Indent for Auto clave given on 22.6.2021 .
5. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- General Maintenance
6. Computer in RVG room not working - General Maintenance on 27.05.2022
7. Amalgamator : Needs pre proportioned capsules - 3 sets
8. Carborundum disc and diamond disc required.

PG

1. Endomotor not working- Given for repair.
2. Pulp tester not working and also need new one.
3. Need 2 apex locator - Requested to Purchase Dept
4. Putty material extruder required
5. IOPAR X ray Machine PCB board to Bangalore for repair on 21.02.2023.
6. RVG/PSP plates for department. Current RVG declared as condemned by satelec sopix company representatives. Requisition given on 13.07.2022
7. Vita 1 line ceramic furnace motherboard complaint not rectified.
8. New X-Ray machine quotation given and approved by Hospital administrator and COO.

Pedodontics

UG

1. 2 Phantom Heads- Installation have to be done.
2. Requesting electrical supply maintenance work of the minor OT adjacent for the purpose of installation of conscious sedation unit and conduct of courses.
3. Model trimmer base has to be fixed to platform/slab it is mobile now.

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement - Requested to Purchase Dept
2. Need one scaler unit in the UG- Requested to Purchase Dept. Urgently required for inspection purpose.





PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai)
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

UG

1. Operator chair - 8Nos required.- Requested to Purchase Dept
2. Double disc model trimmer under maintenance – 1 Disc need to be replaced.
3. Operator chair wheel base broken in chair no. 7 & 10.

Oral Surgery

1. Rhinoplasty kit required- Request given to Purchase Manger on November 2022.
2. Auto clave should be repaired in Operation Theatre.

UG

1. Compressor Air line to be given to all chairs. - waiting for spares.

Oral Medicine

Nil

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel 

Meeting attended by:-

Dr.Palani Prashanth, Department of Prosthodontics 

Dr.John Hearty Deepak, Department of Oral Medicine


Dr.Chrishantha Joybell- Department of Pedodontics 

Dr.Dr.T.JGnanasagar, Department of Oral Surgery

Dr.V.Jevina Christy of Conservative Dentistry 

Dr. Ani Simila, Department of Oral Pathology

Dr.M.Thamilselvan, Department of Periodontics 

Dr.C.Selvakumar, Department Of Public Health Dentistry 

Dr. Bevin Shaga Department of Orthodontics 




PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Date: 22.09.2023

MINUTES OF EQUIPMENT MEETING REPORT

Held in Principal Chamber on 22.09.2023 at 9 AM

Prosthodontics :

1. Steam cleanser nozzle washer needs to be changed. - waiting for spare
2. Digital dewaxing unit needs to be verified from manufacturers. Requested for AMC service.
3. Digital acrylizer unit needs to be verified from manufacturers. Requested for AMC service.
4. Burnout Furnance – i) Operating circuit issue ii) Not closing properly
5. Bego easymix investment vacuum mixer requires replacement of switch board - To contact distributor.

Public Health Dentistry:

1. **Mobile dental unit AC need to be repaired** – General Maintenance.
2. Need one permanent compressor for rural satellite clinic Aralvaimozhi.

Oral Pathology

1. Digital Microscope Display is not Working -Complaint from January 2023
 - Approval pending in February 2023.
 - March 2023 proposal sent & quotation received.
 - April 2023 Mr.Ligesh reported as payment made.
 - May 2023 Maintenance department reported as spares are not available and payment not made.
 - June 2023 Accepted for replacement but still not received.
 - 8th June 2023 taken for repair by Lawrence & Mayo
 - Quotation was requested to company. Email sent to company on 28.06.2023.



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)

KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Quotation sent by company on 01.07.2023.

- Irreparable old display was sent to us by courier.
- For new display quotation has been given to COO on 18.07.2023.
- 14.08.2023 demo for new digital display shown in the department in the presence of COO. Quotation given on 15.08.2023. Reminder Quotation given on 22.09.2023

2. Water Bath Slide warming table to be upgraded.- Requested to Purchase Dept – Requisition given on 28.07.2023
3. Chair no 10, 11 compressor need to repaired – Request for oil free compressor on 28.07.2023.
4. Slides box and slides staining rack needed- Requisition given on 28.07.2023.
5. Monocular light Microscope charge is draining and not working – Requisition for service.
Solutions: 1) Under the slab plug points (provision needed) can be installed.
2) Need to change battery. – URGENT needed for practical exam.
6. Printer is not working, needs to be repaired.

Periodontics

PG :

Nil

UG:

Nil

Conservative Dentistry

UG

1. Need 2 apex locator - Requested to Purchase Dept
2. Required 1 endomotor - Requested to Purchase Dept
3. Automatic developer not working- Need to be condemned
4. Indent for Auto clave given on 22.6.2021 .




PRINCIPAL

Page 2 of 4

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)

KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

5. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- General Maintenance
6. Computer in RVG room not working - General Maintenance on 27.05.2022
7. Amalgamator : Needs pre proportioned capsules - 3 sets
8. Carborundum disc and diamond disc required.

PG

1. Endomotor not working- Given for repair.
2. Pulp tester not working and also need new one.
3. Need 2 apex locator - Requested to Purchase Dept
4. Putty material extruder required
5. IOPAR X ray Machine PCB board sent to Bangalore for repair on 21.02.2023.
6. RVG/PSP plates required for department. Current RVG declared as condemned by satelec sopix company representatives. Requisition given on 13.07.2022
7. Vita 1 line ceramic furnace motherboard complaint not rectified.
8. New X-Ray machine quotation given and approved by Hospital administrator and COO.
9. RVG sensor required before DCI inspection.
10. Putty extruder gun needed.

Pedodontics

UG

1. 2 Phantom Heads- Installation have to be done.
2. Requesting electrical supply maintenance work of the minor OT adjacent to the department for the purpose of installation of conscious sedation unit and conduct of courses.
3. Model trimmer base has to be fixed to platform/slab it is mobile now.

Orthodontics

1. 2 Ultra sonic scaling units required for UG as per DCI requirement - Requested to Purchase Dept
2. Need one scaler unit in the UG- Requested to Purchase Dept. Urgently required for inspection purpose.



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)

KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

1. Operator chair - 8Nos required.- Requested to Purchase Dept
2. Double disc model trimmer under maintenance – 1 Disc need to be replaced.

Oral Surgery

1. Rhinoplasty kit required- Request given to Purchase Manager on November 2022.
2. Autoclave should be repaired in Operation Theatre.

UG

1. Compressor Air line to be given to all chairs. - waiting for spares.

Oral Medicine

Nil

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:-

Dr.Akash Mithran, Department of Prosthodontics

Dr.E.Gracelin Ranjitha, Department of Oral Medicine

Dr.A.Esai Amutha Prabha, Department of Pedodontics

Dr.T.K.Divakar, Department of Oral Surgery

Dr.M.S.Ravisankar, Department of Conservative Dentistry

Dr V.Jevina Christy, Department of Conservative Dentistry

Dr. Ani Simila, Department of Oral Pathology

Dr.P.Fairlin, Department of Periodontics

Dr.P.Ravishankar, Department Of Public Health Dentistry

Dr. Bevin Shaga Department of Orthodontics



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT,



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Date: 27.10.2023

MINUTES OF EQUIPMENT MEETING REPORT

Held in Principal Chamber on 27.10.2023 at 10.15 AM

Prosthodontics :

1. Steam cleanser nozzle washer needs to be changed. - waiting for spare
2. Digital dewaxing unit needs to be verified from manufacturers. Requested for AMC service.
3. Digital acrylizer unit needs to be verified from manufacturers. Requested for AMC service.
4. Burnout Furnance – i) Operating circuit issue ii) Not closing properly
5. Bego easymix investment vacuum mixer requires replacement of switch board - To contact distributor.

Public Health Dentistry:

1. Mobile dental unit AC need to be repaired – General Maintenance.

Oral Pathology

1. Digital Microscope Display is not Working - Complaint from January 2023
 - Approval pending in February 2023.
 - March 2023 proposal sent & quotation received.
 - April 2023 Mr.Ligesh reported as payment made.
 - May 2023 Maintenance department reported as spares are not available and payment not made.
 - June 2023 Accepted for replacement but still not received.
 - 8th June 2023 taken for repair by Lawrence & Mayo
 - Quotation was requested to company. Email sent to company on 28.06.2023.



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Quotation sent by company on 01.07.2023.

- Irreparable old display was sent to us by courier.
- For new display quotation has been given to COO on 18.07.2023.
- 14.08.2023 demo for new digital display shown in the department in the presence of COO. Quotation given on 15.08.2023. Reminder Quotation given on 22.09.2023 via E-Mail.

2. Water Bath Slide warming table to be upgraded.- Requested to Purchase Dept – Requisition given on 28.07.2023
3. Chair no 10, 11 compressor need to repaired – Request for oil free compressor on 28.07.2023.
4. Slides box and slides staining rack needed- Requisition given on 28.07.2023.
5. Monocular light Microscope charge is draining and not working – Requisition for service. Need to change battery for 10 nos. – URGENT needed for practical exam.

Periodontics

PG :

Nil

UG:

Nil

Conservative Dentistry


UG

1. Need 2 apex locator - Requested to Purchase Dept
2. Required 1 endomotor - Requested to Purchase Dept
3. Automatic developer not working- Need to be condemned
4. Indent for B CLASS Autoclave given on 22.6.2021 .
5. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- General Maintenance
6. Computer in RVG room not working - General Maintenance on 27.05.2022


PRINCIPAL

Page 2 of 4

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.





RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

7. Amalgamator : Needs pre proportioned capsules - 3 sets
8. Carborundum disc and diamond disc required.

PG

1. Endomotor not working- Given for repair.
2. Pulp tester not working and also need new one.
3. Need 2 apex locator - Requested to Purchase Dept
4. Putty material extruder required
5. IOPAR X ray Machine PCB board sent to Bangalore for repair on 21.02.2023 – Company said it cannot be repaired
6. RVG/PSP plates required for department. Current RVG declared as condemned by satelec sopix company representatives. Requisition given on 13.07.2022
7. Vita 1 line ceramic furnace motherboard complaint not rectified.
8. New X-Ray machine quotation given and approved by Hospital administrator and COO.
9. RVG sensor required before DCI inspection.
10. Putty extruder gun needed.
11. Instruments and Equipments needed for 3 unit division- Requisition given on 03.10.2023.

Pedodontics

UG

1. 2 Phantom Heads- Installation have to be done.
2. Requesting electrical supply maintenance work of the minor OT adjacent to the department for the purpose of installation of conscious sedation unit and conduct of courses.
3. Model trimmer base has to be fixed to platform/slab it is mobile now.

Orthodontics

1. 2 Ultra sonic scaling units required for UG as per DCI requirement - Requested to Purchase Dept
2. Need one scaler unit in the UG- Requested to Purchase Dept. Urgently required for inspection purpose.

UG

1. Operator chair - 8Nos required.- Requested to Purchase Dept




PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL

(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai Accredited by NAAC)

KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

2. Double disc model trimmer under maintenance – 2 Disc need to be replaced.

Oral Surgery

1. Rhinoplasty kit required- Request given to Purchase Manager on November 2022.
2. Autoclave should be repaired in Operation Theatre.

UG

1. Compressor Air line to be given to all chairs. - waiting for spares.

Oral Medicine

Nil

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:-

Dr.Akash Mithran, Department of Prosthodontics

Dr.A.Nelson, Department of Oral Medicine

Dr.Shana Yasmin, Department of Pedodontics

Dr.N.T.Jenin, Department of Oral Surgery

Dr.K.T.Manoj, Department of Conservative Dentistry

Dr.C.S.Ani Simila, Department of Oral Pathology

Dr.M.Thamilselvan, Department of Periodontics

Dr.P.Ravishankar, Department Of Public Health Dentistry

Dr.I.Bevin Shaga Department of Orthodontics




PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Date: 28.07.2023

MINUTES OF EQUIPMENT MEETING REPORT

Held in Principal Chamber on 28.07.2023 at 9.00 A.M

Prosthodontics :

1. Steam cleanser nozzle washer needs to be changed. - waiting for spare
2. Digital dewaxing unit needs to be verified from manufacturers. Requested for AMC service.
3. Digital acrylizer unit needs to be verified from manufacturers. Requested for AMC service.
4. Burnout Furnance – i) Operating circuit issue ii) Not closing properly
5. Ray Foster spindle grinder second unit also not working (total of 2 units only). Both are non functional
6. UG clinic: staff table repair - MRL given on 21.07.2023

Public Health Dentistry:

1. Mobile dental unit AC need to be repaired – General Maintenance
2. Compressor-1hp need to be installed in Aralvaimozhi rural satellite clinic

Oral Pathology

1. Digital Microscope Display is not Working -Complaint from January 2023
 - Approval pending in February 2023.
 - March 2023 proposal sent & quotation received.
 - April 2023 Mr.Ligesh reported as payment made.
 - May 2023 Maintenance department reported as spares are not available and payment not made.
 - June 2023 Accepted for replacement but still not received.
 - 8th June 2023 taken for repair by Lawrence & Mayo



PRINCIPAL
RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai)
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

- Quotation was requested to company. Email sent to company on 28.06.2023. Quotation sent by company on 01.07.2023.
 - Irreparable old display was sent to us by courier.
 - For new display quotation has been given to COO on 18.07.2023
2. Water Bath Slide warming table to be upgraded.- Requested to Purchase Dept
 3. Chair no 10, 11 compressor need to repaired – Request for oil free compressor on 28.07.2023.
 4. Slides box and slides staining rack needed.
 5. Monocular light Microscope charge is draining and not working.
Solutions: 1) Under the slab plug points (provision needed) can be installed.
2) Need to change battery. – URGENT needed for practical exam.

Periodontics


PG & UG

Nil

Conservative Dentistry

UG

1. Need 2 apex locator - Requested to Purchase Dept
2. Required 1 endomotor - Requested to Purchase Dept
3. Automatic developer not working- Need to be condemned
4. Indent for Autoclave given on 22.6.2021 .
5. Replacement of cupboards in Sterilization room - indent given on 22.04.2021- General Maintenance
6. Computer in RVG room not working - General Maintenance on 27.05.2022
7. Amalgamator : Needs pre proportioned capsules - 3 sets
8. Carborundum disc and diamond disc required.



PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

PG

1. Endomotor not working- Given for repair.
2. Pulp tester not working and also need new one.
3. Need 2 apex locator - Requested to Purchase Dept
4. Putty material extruder required
5. IOPAR X ray Machine PCB board to Bangalore for repair on 21.02.2023.
6. RVG/PSP plates for department. Current RVG declared as condemned by satelec sopix company representatives. Requisition given on 13.07.2022
7. Vita 1 line ceramic furnace motherboard complaint not rectified.
8. New X-Ray machine quotation given and approved by Hospital administrator and COO.
9. Pre clinical central phantom lab – Compressor not connected.

Pedodontics

UG

1. 2 Phantom Heads which have been shifted from Conservative department need to be connected and made fully functional.
2. Requesting electrical supply maintenance work of the minor OT adjacent for the purpose of installation of conscious sedation unit and conduct of courses.

Orthodontics

1. 2 Ultra sonic cleaner required for UG as per DCI requirement - Requested to Purchase Dept
2. Need one scaler unit in the UG- Requested to Purchase Dept. Urgently required for inspection purpose.

UG

1. Operator chair - 8Nos required.- Requested to Purchase Dept
2. Double disc model trimmer under maintenance – 1 Disc need to be replaced





PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT,



RAJAS DENTAL COLLEGE & HOSPITAL
(Affiliated to The Tamil Nadu Dr. M.G.R Medical University, Chennai
Accredited by NAAC)
KAVALKINARU JN., TIRUNELVELI DISTRICT – 627105

Oral Surgery

1. Rhinoplasty kit required- Request given to Purchase Manger on November 2022.
2. Autoclave should be repaired in Operation Theatre.

UG


1. Compressor Airline to be given to all chairs. - waiting for spares.


Oral Medicine

1. IOPA machine is not functioning – Alerio DC optoma mobile unit – Given for repair on 22.07.2023
2. Acquisition of PSP radiographic units.
3. 0.75 or 0.8 compressor or dual head 0.8 compressor for speciality clinic quotation are sent.

Meeting Chaired by: - Principal Prof. Dr. Alex Mathews Muruppel

Meeting attended by:-


Dr.Sarath Chandra Govindraj, Department of Prosthodontics 

Dr.Palani Prashanth, Department of Prosthodontics 

Dr.John Hearty Deepak, Department of Oral Medicine 

Dr.Chrishantha Joybell- Department of Pedodontics 

Dr.Divakar, Department of Oral Surgery

Dr.V.Jevina Christy of Conservative Dentistry 

Dr. Ani Simila, Department of Oral Pathology

Dr.M.Thamilselvan, Department of Periodontics 

Dr.C.Selvakumar, Department Of Public Health Dentistry 

Dr. Bevin Shaga Department of Orthodontics 




PRINCIPAL

RAJAS DENTAL COLLEGE & HOSPITAL
KAVALKINARU JN - 627 105.
TIRUNELVELI DISTRICT.